

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Sterile Transport Swab (Swab w/Tube) 5"	
Product Code: PW047-100PC	Expiry Date: 15/01/2021
Lot Number: 00010116	Release Date: 15/01/2016
Analytical Report Number: QAD/AR-16/02-236	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATIONS	OBSERVATIONS	REMARKS
Appearance	Transparent, clear & unbreakable, Polypropylene tube with specially designed cap and cotton bud with polypropylene stick. Individually packed, Size: 125mm x 14mm diameter. Pack of 100 nos.	Transparent, clear & unbreakable, Polypropylene tube with specially designed cap and cotton bud with polypropylene stick. Individually packed, Size: 125mm x 14mm diameter. Pack of 100 nos.	Complies
Sterilization	The above batch of Petri plates was sterilized by Microtrol Sterilization Services Pvt. Ltd. by treatment with Ethylene oxide to the following parameters: Concentration of Ethylene oxide : NLT 750 gm/M3 Exposure time : 6 hours Humidity : Min. 50% RH Temperature : 50 Deg C	Concentration of Ethylene oxide: 750 gm/M3	Complies
Sterility Testing	As per USP36/NF31.	Sterility passes	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4780, 4681, 9320126789
Email: micromaster@micromasterlab.com

