

# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: T.A.T. Broth (Pack of 10 X 100ml)	
Product Code: RM098	Expiry Date: 18/01/2020
Lot Number: 00390116	Release Date: 18/01/2016
Analytical Report Number: QAD/AR-16/01-100	Page No: 1 of 2

PREPARED MEDIA			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Slightly opalescent solution	Slightly opalescent solution	Complies
Color of Final Medium	Light yellow colored medium	Light yellow colored medium	Complies
Clarity of Final Medium	Slightly opalescent solution in bottle	Slightly opalescent solution in bottle	Complies
pH at 25°C	7.00-7.40	7.10	Complies
Gel Strength	Not applicable	Not applicable	Complies
Quantity of Medium	100 ml	100ml	Complies
PREINCUBATION CHECK			
Pre Incubation Check	No growth should be observed after incubation at 30-35 °C for 24 - 48 hrs	No growth observed after incubation at 30-35 °C for 24 - 48 hrs	Complies

## CULTURAL PARAMETERS: Cultural chara

Organisms	Results to be achieved		
	Inoculum (CFU)	Growth	Remark
<i>Bacillus subtilis</i> ATCC 6633	77	Good-luxuriant	Complies
<i>Candida albican</i> ATCC10231	94	Good-luxuriant	Complies
<i>Pseudomonas aeruginosa</i> ATCC27853	92	Good-luxuriant	Complies
<i>Salmonella Typhi</i> ATCC 6539	82	Good-luxuriant	Complies
<i>Staphylococcus aureus</i> ATCC25923	79	Good-luxuriant	Complies

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126789  
Email: micromaster@micromasterlab.com



# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY






Product Name: T.A.T. Broth (Pack of 10 X 100ml)	
Product Code: RM098	Expiry Date: 18/01/2020
Lot Number: 00390116	Release Date: 18/01/2016
Analytical Report Number: QAD/AR-16/01-100	Page No: 2 of 2

Organisms	Results to be achieved		
	Inoculum (CFU)	Growth	Remark
<i>Staphylococcus aureus</i> ATCC6538	82	Good-luxuriant	Complies
<i>Pseudomonas aeruginosa</i> 9027	82	Good-luxuriant	Complies

## Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

Culturing your needs in Microbiology.

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126789  
Email: micromaster@micromasterlab.com

