

# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Penicillin Discs (Symbol: P or PG; Potency: 10 units)	
Product Code: AS039	Expiry Date: 29/07/2019
Lot Number: 00640116	Release Date: 29/01/2016
Analytical Report Number: QAD/AR-16/01-154	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "P or PG" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "P or PG" on each side of the disc.	Complies

### Cultural Parameters:

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter in mm	Zone Diameter Observed in mm	Remarks
1	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	26-37	29	Complies
2	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	24-30	27	Complies
3	<i>Neisseria gonorrhoeae</i> ATCC 49226	GC Agar Base with 1% GC supplement (MS016)	35-37°C/ 18-24 hrs in 5-10% CO <sub>2</sub> + 70% humidity	26-34	28	Complies

### Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126789  
Email: micromaster@micromasterlab.com

