

# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Norfloxacin Discs (Symbol: NO or NX ; Potency: 10 mcg)	
Product Code: AS035-1VI	Expiry Date: 03/08/2019
Lot Number: 00660116	Release Date: 03/02/2016
Analytical Report Number: QAD/AR-16/01-160	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "NO" or "NX" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "NO" or "NX" on each side of the disc.	Complies

## Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar ( DM172)	32-35°C/ 18-24 hrs	28-35	35	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar DM172)	32-35°C/ 18-24 hrs	17-28	27	Complies
3	<i>Pseudomonas aeruginosa</i> ATCC 27853	Mueller Hinton Agar ( DM172)	32-35°C/ 18-24 hrs	22-29	28	Complies
4	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	15-21	18	Complies

## Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

<b>TESTED BY</b> (Sneha Daragshetti) MICROBIOLOGIST	<b>CHECKED BY</b> (Swati More) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Result

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4881, 9320126789  
Email: micromaster@micromasterlab.com

