

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Grams Iodine	
Product Code: SI016-100ml	Expiry Date : 30/01/2024
Lot Number: 00920216	Release Date: 30/01/2016
Analytical Report Number: QAD/AR-16/02-224	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Reddish Brown colored solution	Reddish Brown colored solution	Complies
Staining of Blood Smear	Gram Staining procedure followed using the reagent should show appropriate results	Gram Staining procedure followed using the reagent showed appropriate results	Complies
Quality Control strains used	<i>Staphylococcus aureus</i> ATCC 25923	Gram-positive cocci	Complies
	<i>Escherichia coli</i> ATCC 25922	Gram-negative rods	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology

38/39, Kalpetaru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Rurwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

