

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Master Loop w/Holder-3	
Product Code: ML003	Expiry Date: 05/02/2024
Lot Number: 00980216	Release Date: 05/02/2016
Analytical Report Number: QAD/AR-16/02-254	Page No: 1 of 1

Environmental Detail: Temperature: 25.6°C

Humidity: 51%

Test Conducted	Master Loop Diameter(mm)	Master Loop Volume(μ L)	Expected Uncertainty
Reading 1	3.0	7.0	+/- 0.02
Reading 2	3.1	7.1	
Reading 3	3.1	7.1	
Reading 4	2.1	7.1	
Reading 5	3.0	7.2	
Reading 6	3.0	7.1	
Reading 7	3.0	7.0	
Reading 8	3.0	7.0	
Reading 9	3.2	7.0	
Reading 10	3.0	7.1	
Average	3.1	7.1	

Traceability:

Instrument Details:		
Name:	ID No	Calibration Due Date
Digital Weighing Balance	MAC-034	---
Weight Box	MI-032	04/08/2016

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00 Micromaster Laboratories Pvt. Ltd. 10/12/2014
Manufacture & Exporter

QAD/FR/01
Culturing your needs in Microbiology.

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

