

# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY






Product Name: X-Factor Discs	
Product Code: ID007-1vl	Expiry Date : 22/08/2019
Lot Number: 01090216	Release Date: 22/02/2016
Analytical Report Number: QAD/AR-16/02-359	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Description	Sterile filter paper discs, impregnated with a solution of hemin reagent	Sterile filter paper discs, impregnated with a solution of hemin reagent	Complies
X-Factor Discs	<i>Haemophilus influenzae</i> ATCC 49766	No zone of growth around the disc	Complies
	<i>Haemophilus parainfluenzae</i> ATCC 33392	No zone of growth around the disc	
	<i>Haemophilus ducreyi</i> ATCC 700724	A clearly defined zone of growth around the disc	
	<i>Bordetella pertussis</i> ATCC 9797	A clearly defined zone of growth around the disc	

## Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Sneha Daragshetti) MICROBIOLOGIST	<b>CHECKED BY</b> (Swati More) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126789  
Email: micromaster@micromasterlab.com

