

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: ONPG Discs	
Product Code: ID031	Expiry Date : 19/08/2019
Lot Number: 01100216	Release Date: 19/02/2016
Analytical Report Number: QAD/AR-16/02-339	Page No: 1 of 2

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 10 mm diameter with printed "OX" on each side of the disc.	Filter paper discs of 10 mm diameter with printed "OX" on each side of the disc.	Complies

Cultural Parameters

ONPG reaction observed in 0.85% sodium chloride solution of following culture containing ONPG disc after an incubation of upto 4 hours at 35-37°C.

Sr. No.	Test Organisms	Incubation time	Observation	Remarks
1.	<i>Escherichia coli</i> ATCC 25922	After 4 hours at 35-37°C	Positive reaction; yellow color	Complies
2.	<i>Citrobacter freundii</i> ATCC8090	After 4 hours at 35-37°C	Positive reaction; yellow color	Complies
3.	<i>Enterobacter aerogenes</i> ATCC 13048	After 4 hours at 35-37°C	Positive reaction; yellow color	Complies
4.	<i>Salmonella Typhimurium</i> ATCC 14028	After 4 hours at 35-37°C	Negative reaction; no color change	Complies
5.	<i>Salmonella Choleraesuis</i> ATCC 12011	After 4 hours at 35-37°C	Positive reaction; yellow color	Complies
6.	<i>Proteus vulgaris</i> ATCC 13315	After 4 hours at 35-37°C	Negative reaction; no color change	Complies

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacturer & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: ONPG Discs	
Product Code: ID031	Expiry Date : 19/08/2019
Lot Number: 01100216	Release Date: 19/02/2016
Analytical Report Number: QAD/AR-16/02-339	Page No: 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kelpetaru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

