



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Bacitracin Discs (Symbol: B ; Potency: 10 units)	
Product Code: AS006	Expiry Date: 13/08/2019
Lot Number: 01110216	Release Date: 13/02/2016
Analytical Report Number: QAD/AR-16/02-292	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "B" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "B" on each side of the disc.	Complies

Cultural Parameters:

Sr. No.	Test Organisms	Media used (code number)	Incubation time / temperature	Standard Zone Diameter	Zone Diameter Observed in mm	Remarks
1	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	12-22	22	Complies

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

