

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Streptomycin Discs (Symbol: ST or S; Potency: 10mcg)	
Product Code: AS043-1vl	Expiry Date: 18/08/2019
Lot Number: 01120216	Release Date: 18/02/2016
Analytical Report Number: QAD/AR-16/02-323	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "ST or S " on each side of the disc.	Filter paper discs of 6 mm diameter with printed "ST or S " on each side of the disc.	Complies

Cultural Parameters:

Sr. No.	Test Organisms	Media used (code number)	Incubation time / temperature	Standard Zone Diameter	Zone Diameter Observed in mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	12-20	18	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	14-22	16	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Soniya Sebastian) MICROBIOLOGIST	CHECKED BY (Vijayalaxmi Pujari) HEAD MICROBIOLOGY	APPROVED BY (Swati More) HEAD QUALITY ASSURANCE

End of the Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Karpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

