

# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Vancomycin Discs (Symbol: VA; Potency: 30 mcg)	
Product Code: AS047	Expiry Date: 24/08/2019
Lot Number: 01370216	Release Date: 24/02/2016
Analytical Report Number: QAD/AR-16/02-355	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filler paper discs of 6 mm diameter with printed "VA" on each side of the disc.	Filler paper discs of 6 mm diameter with printed "VA" on each side of the disc.	Complies

## Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	17-21	21	Complies
2	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	20-27	23	Complies

## Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126789  
Email: micromaster@micromasterlab.com

