



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Yeast Extract Agar	
Product Code: DM298	Expiry Date: 05/03/2021
Lot Number: 01540216	Release Date: 05/03/2016
Analytical Report Number: QAD/AR-16/03-459	Page No: 1 of 2

DEHYDRATED MEDIUM

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color	Cream to yellow colored powder	Cream to yellow colored powder	Complies
Consistency	Homogenous, free flowing powder	Homogenous, free flowing powder	Complies

PREPARED MEDIUM

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color of medium	Yellow colored Solution	Yellow colored Solution	Complies
Clarity of medium (After sterilization)	Clear to slightly opalescent gel forms in Petri plates.	Clear to slightly opalescent gel forms in Petri plates.	Complies
pH (After sterilization)	7.00-7.40 at 25°C	7.00	Complies
Gel strength	Firm as compared to 1.5% Agar	Firm as compared to 1.5% Agar	Complies

Rev. 00


10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com





CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY




Product Name: Yeast Extract Agar	
Product Code: DM298	Expiry Date: 05/03/2021
Lot Number: 01540216	Release Date: 05/03/2016
Analytical Report Number: QAD/AR-16/03-459	Page No: 2 of 2

CULTURAL PARAMETERS:

Organisms	Incubation time/ temp	No. of organisms inoculated	No. of organisms recovered	Percentage Recovery	Growth	Remarks
<i>Enterobacter aerogenes</i> ATCC 13048	35-37°C / 18-24 hrs.	75	59	79%	Luxuriant	Complies
<i>Escherichia coli</i> ATCC 25922	35-37°C / 18-24 hrs.	87	65	75%	Luxuriant	Complies
<i>Pseudomonas aeruginosa</i> ATCC 27853	35-37°C / 18-24 hrs.	90	70	78%	Luxuriant	Complies
<i>Staphylococcus aureus</i> ATCC 25923	35-37°C / 18-24 hrs.	75	58	77%	Luxuriant	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Soniya Sebastian) MICROBIOLOGIST	CHECKED BY (Vijayalaxmi Pujari) HEAD MICROBIOLOGY	APPROVED BY (Swati More) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

