

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Sparfloxacin Discs (Symbol: SPX or SP; Potency: 5mcg)	
Product Code: AS056	Expiry Date: 02/09/2019
Lot Number: 01600216	Release Date: 02/03/2016
Analytical Report Number: QAD/AR-16/02-406	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "SPX or SP " on each side of the disc.	Filter paper discs of 6 mm diameter with printed " SPX or SP " on each side of the disc.	Complies

Cultural Parameters:

Sr. No.	Test Organisms	Media used (code number)	Incubation time / temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	30-38	35	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	27-33	30	Complies
3	<i>Pseudomonas aeruginosa</i> ATCC 27853	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	21-29	24	Complies
4	<i>Neisseria gonorrhoeae</i> ATCC 49226	GC Agar Base with 1% GC supplement (MS016)	32-35°C/ 40-48 hrs	43-51	45	Complies
5	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	21-27	24	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology.

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4881, 9320126789
Email: micromaster@micromasterlab.com

