

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Amoxicillin +Clavulanic acid Discs (Symbol: AMC or AU; Potency: 20+10mcg)	
Product Code: AS004	Expiry Date: 02/09/2019
Lot Number: 01620216	Release Date: 02/03/3016
Analytical Report Number: QAD/AR-16/02-408	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 5 mm diameter with printed "AMC or AU" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "AMC or AU" on each side of the disc.	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	18-24	24	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	28-36	35	Complies
3	<i>Haemophilus influenzae</i> ATCC 49247	Haemophilus Test Agar Base(DM1273) with added Haemophilus Growth Supplement (MS085)	35-37°C/ 18-24 hrs in 5-7% CO ₂	15-23	17	Complies
4	<i>Escherichia coli</i> ATCC 35218	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	17-22	18	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

