

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Sterile Disposable Plates-90mm	
Product Code: PWD12	Expiry Date: 24/02/2021
Lot Number: 01650216	Release Date: 24/02/2016
Analytical Report Number: QAD/AR-16/02-368	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATIONS	OBSERVATIONS	REMARKS
Appearance	Optically clear, Polystyrene, Petri plates, double packed in stacks of 10. Size: 90mm diameter x 15mm	Optically clear, Polystyrene, Petri plates, double packed in stacks of 10. Size: 90mm diameter x 15mm	Complies
Sterilization	The above batch of Petri plates was sterilized by Microtrol Sterilization Services Pvt. Ltd. by treatment with Ethylene oxide to the following parameters: Concentration of Ethylene oxide : NLT 750 gm/M3 Exposure time : 6 hours Humidity : Min. 50% RH Temperature : 50 Deg C	Concentration of Ethylene oxide: 750 gm/M3	Complies
Sterility Testing	As per USP36/NF31.	Sterility passes	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology.

38/39, Kalpetaru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

