

# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Master Loop w/Holder-4	
Product Code: ML004	Expiry Date: 02/03/2024
Lot Number: 01830316	Release Date: 02/03/2016
Analytical Report Number: QAD/AR-16/03-428	Page No: 1 of 1

Environmental Detail: Temperature: 25.6°C

Humidity: 51%

Test Conducted	Master Loop Diameter(mm)	Master Loop Volume( $\mu$ L)	Expected Uncertainty
Reading 1	4.1	10.1	+/- 0.02
Reading 2	4.0	10.1	
Reading 3	4.0	10.0	
Reading 4	4.0	10.0	
Reading 5	4.1	10.0	
Reading 6	3.9	10.1	
Reading 7	4.1	10.1	
Reading 8	4.1	10.1	
Reading 9	4.0	10.0	
Reading 10	4.0	10.1	
Average	4.0	10.1	

## Traceability:

Instrument Details:		
Name:	ID No	Calibration Due Date
Digital Weighing Balance	MAC-034	---
Weight Box	MI-032	04/08/2016

## Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

<b>TESTED BY</b> (Sneha Daragshetti) MICROBIOLOGIST	<b>CHECKED BY</b> (Swati More) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

10/12/2014

QAD/FR/01

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126760  
Email: micromaster@micromasterlab.com

