



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Master Anaerobic GasPak (For Anaerobic Jar of 3.5L Capacity)	
Product Code: AAN001-1BX	Expiry Date: 10/03/2018
Lot Number: 01970316	Release Date: 10/03/2016
Analytical Report Number: QAD/AR-16/03-468	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Description	GasPak for generation of Hydrogen and Carbon Dioxide in Anaerobic Jars of 3.5L Capacity	GasPak for generation of Hydrogen and Carbon Dioxide in Anaerobic Jars of 3.5L Capacity	Complies
Cultural Parameters	Cultural characteristics observed using Master Anaerobic GasPak in an anaerobic atmosphere, after incubation at should be as shown in the Table below	Cultural characteristics observed using Master Anaerobic GasPak in an anaerobic atmosphere, after incubation at 30-35°C for 24-48 hours should be as shown in the Table below	Complies

CULTURAL PARAMETERS:

Organisms	Incubation temp/ time	Growth	Remarks
<i>Clostridium sporogenes</i> ATCC 19404	30-35°C / 24-48 hrs	Good-Luxuriant	Complies
<i>Clostridium perfringens</i> ATCC 13124	30-35°C / 24-48 hrs	Good-Luxuriant	Complies
<i>Pseudomonas aeruginosa</i> ATCC 10145	30-35°C / 24-48 hrs	Poor growth – No growth	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

