

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: MacConkey Agar w/o CV, w/o NaCl	
Product Code: DM1557	Expiry Date : 11/03/2021
Lot Number: 02110316	Release Date: 11/03/2016
Analytical Report Number: QAD/AR-16/03-474	Page No: 1 of 2

DEHYDRATED MEDIUM			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color	Light yellow to light pink colored powder	Light yellow to light pink colored powder	Complies
Consistency	Homogenous, free flowing powder	Homogenous, free flowing powder	Complies
PREPARED MEDIUM			
TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Color of medium	Orange red colored medium	Orange red colored medium	Complies
Clarity of medium (After sterilization)	Clear to slightly opalescent gel forms in Petri plates	Clear to slightly opalescent gel formed in Petri plates	Complies
pH (After sterilization)	6.90-7.30 at 25°C	7.00	Complies
Gel strength	Firm as compared to 1.2% Agar gel	Firm as compared to 1.2% Agar gel	Complies

CULTURAL PARAMETERS: Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours.

Organisms	Incubation time/ temp	No. of organisms inoculated	No. of organisms recovered	Percentage recovery	Growth	Color of colony	Remarks
<i>Escherichia coli</i> ATCC 25922	35 - 37°C / 18 - 24 hrs.	87	68	78%	Good-luxuriant	Pink to red with Bile precipitate	Complies
<i>Enterobacter aerogenes</i> ATCC 13048	35 - 37°C / 18 - 24 hrs.	86	65	76%	Good-luxuriant	Pink to red	Complies
<i>Enterococcus faecalis</i> ATCC 29212	35 - 37°C / 18 - 24 hrs.	74	27	36%	Fair to good	Colorless to pale pink	Complies
<i>Proteus vulgaris</i> ATCC 13315	35 - 37°C / 18 - 24 hrs.	66	49	74%	Good-luxuriant	Colorless	Complies
<i>Salmonella paratyphi A</i> ATCC 9150	35 - 37°C / 18 - 24 hrs.	72	56	78%	Good-luxuriant	Colorless	Complies
<i>Shigella flexneri</i> ATCC 12022	35 - 37°C / 18 - 24 hrs.	63	24	38%	Fair to good	Colorless	Complies
<i>Salmonella paratyphi B</i> ATCC 8759	35 - 37°C / 18 - 24 hrs.	68	51	75%	Good-luxuriant	Colorless	Complies

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

39/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2599 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY






Product Name: MacConkey Agar w/o CV, w/o NaCl	
Product Code: DM1557	Expiry Date : 11/03/2021
Lot Number: 02110316	Release Date: 11/03/2016
Analytical Report Number: QAD/AR-16/03-474	Page No: 2 of 2

CULTURAL PARAMETERS:

Organisms	Incubation time/ temp	No. of organisms inoculated	No. of organisms recovered	Percentage recovery	Growth	Color of colony	Remarks
<i>Salmonella enteritidis</i> ATCC 13076	35 - 37°C / 18 - 24 hrs.	74	55	74%	Good-luxuriant	Colorless	Complies
<i>Salmonella typhi</i> ATCC 6539	35 - 37°C / 18 - 24 hrs.	72	52	72%	Good-luxuriant	Colorless	Complies
<i>Staphylococcus aureus</i> ATCC 25923	35 - 37°C / 18 - 24 hrs.	$\geq 10^3$	0	0%	Inhibited	--	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Shubhangi Kachare) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

