

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Sucrose (Su)	
Product Code: ID026	Expiry Date : 21/10/2019
Lot Number: 02250316	Release Date: 21/04/2016
Analytical Report Number: QAD/AR-16/04-603	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 10 mm diameter , printed with letters "Su" on both sides of the discs	Filter paper discs of 10 mm diameter , printed with letters "Su" on both sides of the discs	Complies




CULTURAL PARAMETERS:

The carbohydrate fermentation reactions after an incubation of various bacteria with Dextrose Differentiation discs were tested using Andrade's Peptone water (DM 013).

Organisms	Incubation time/ temp	Growth	Acid	Gas	Remarks
<i>Escherichia coli</i> ATCC 25922	18-48 hrs / 35-37°C	Luxuriant	Positive reaction: Pink color	Positive reaction	Complies
<i>Proteus mirabilis</i> ATCC12453	18-48 hrs / 35-37°C	Luxuriant	Negative reaction: no color change	Negative reaction	Complies
<i>Proteus vulgaris</i> ATCC 8427	18-48 hrs / 35-37°C	Luxuriant	Positive reaction: Pink color	Positive reaction	Complies
<i>Salmonella Typhimurium</i> ATCC 14028	18-48 hrs / 35-37°C	Luxuriant	Negative reaction: no color change	Negative reaction	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Soniya Sebastian) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

