

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Oxacillin Discs OX 1mcg (Symbol: OX; Potency: 1 mcg)	
Product Code: AS091	Expiry Date: 23/09/2019
Lot Number: 02270316	Release Date: 23/03/2016
Analytical Report Number: QAD/AR-16/03-530	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "OX" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "OX" on each side of the disc.	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used with (code number)	Incubation time and temperature	Standard Zone Diameter in mm	Zone Diameter Observed in mm	Remarks
1	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	18-24	24	Complies
2	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	≤12	11	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Sneha Daragshetti) MICROBIOLOGIST	CHECKED BY (Swati More) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

36/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlabs.com

