

# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY






Product Name: Sterile Flexi Loops 2.2 mm	
Product Code: PW035	Expiry Date: 09/04/2021
Lot Number: 02450316	Release Date: 09/04/2016
Analytical Report Number: QAD/AR-16/04-619	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATIONS	OBSERVATIONS	REMARKS
Appearance	Sterilized Flexi Loops, 2.2 mm diameter, calibrated to 0.005ml, Individually sealed, Pack of 100 / 500 loops.	Sterilized Flexi Loops, 2.2 mm diameter, calibrated to 0.005ml, Individually sealed, Pack of 100 / 500 loops.	Complies
Sterilization	The above batch of Flexi Loops was sterilized by Agrosurge Irradiators Pvt. Ltd. by treatment with Gamma Irradiation and dose should be given 15Kgy.	Dose of Gamma Irradiation:15Kgy	Complies
Sterility Testing	As per USP36/NF31.	Sterility passes	Complies

## Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Soniya Sebastian) MICROBIOLOGIST	<b>CHECKED BY</b> (Swati More) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Rev. 00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwel Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126789  
Email: micromaster@micromasterlab.com

