# CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

| | |
|---|---|
| **Product Name: Sterile Flexi Loops 2.2 mm** | |
| **Product Code: PW035-500 PC** | **Expiry Date: 23.12.2020** |
| **Lot Number: 02911215** | **Release Date: 23.12.2015** |
| **Analytical Report Number: QAD/AR-15/12-357** | **Page No: 1 of 1** |

| TEST CONDUCTED | SPECIFICATIONS | OBSERVATIONS | REMARKS |
|---|---|---|---|
| Appearance | Sterilized Flexi Loops, 2.2 mm diameter, calibrated to 0.005ml, Individually sealed, Pack of 100 / 500 loops. | Sterilized Flexi Loops, 2.2 mm diameter, calibrated to 0.005ml, Individually sealed, Pack of 100 / 500 loops. | Complies |
| Sterilization | The above batch of Flexi Loops was sterilized by Agrosurge Irradiators Pvt. Ltd. by treatment with Gamma Irradiation and dose should be given 15Kgy. | Dose of Gamma Irradiation:15Kgy | Complies |
| Sterility Testing | As per USP36/NF31. | Sterility passes | Complies |

## Final Results: COMPLIES

**This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.**

| | | |
|---|---|---|
| **TESTED BY (Shubhangi Kachare) MICROBIOLOGIST** | **CHECKED BY (Swati More) HEAD MICROBIOLOGY** | **APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE** |

**End Of The Report**



**Micromaster Laboratories Pvt. Ltd.**
Manufacture & Exporter
Culturing your needs in Microbiology...
38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd., Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com