CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Antimicrobial susceptibility multiple disc (12-in-1 Combi Discs)	
Product Code: MD012S	Expiry Date: 03/12/2019
Lot Number: 03580516	Release Date: 03/06/2016
Analytical Report Number: QAD/AR-16/06-764	Page No: 1 of 2

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper rings of 12 discs of 6 mm diameter with printed "SYMBOL" on one side of the rings.	Filter paper rings of 12 discs of 6 mm diameter with printed "SYMBOL" on one side of the rings.	Complies
TEST CONDITION			
Media used (Product Code)	Mueller Hinton Agar(DM172)	Mueller Hinton Agar(DM172)	Complies
Incubation temperature and time	32-35°C/18-24 hrs	32-35°C/18-24 hrs	Complies

Cultural Parameters

Sr. No.	Antibiotics Used	Standard Zone Diameter	Zone Diameter Observed In mm	Standard Zone Diameter	Zone Diameter Observed In mm	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
		Escherichia coli ATCC 25922		*Staphylococcus aureus ATCC 25923*		*Pseudomonas aeruginosa ATCC 27853*		
1.	Ceftriazone A1-30mcg	29-35	30	22-28	28	17-23	18	Complies
2.	Amikacin A2-30mcg	19-26	30	20-26	22	18-26	18	Complies
3.	Imipenem A3-10mcg	26-32	27	--	--	20-28	20	Complies
4.	Ofloxacin A4-5mcg	29-33	30	24-28	28	17-21	21	Complies
5.	Cefuroxime A5-30mcg	20-26	25	27-35	35	--	--	Complies
6.	Levofloxacin A6 -5mcg	29-37	32	25-30	25	19-26	24	Complies
7.	Cefolaxime A7-30mcg	29-35	29	25-31	26	18-22	20	Complies

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Antimicrobial susceptibility multiple disc (12-in-1 Combi Discs)	
Product Code: MD012S	Expiry Date: 03/12/2019
Lot Number: 03580516	Release Date: 03/06/2016
Analytical Report Number: QAD/AR-16/06-764	Page No: 2 of 2

8.	Ceftazidime A8-30mcg	25-32	32	16-20	20	22-29	29	Complies
9.	Meropenem A9-10mcg	28-34	32	29-37	37	27-33	32	Complies
10.	Gentamicin A10-10mcg	19-26	24	19-27	26	16-21	21	Complies
11.	Ciprofloxacin A11-5 mcg	30-40	32	22-30	30	25-33	32	Complies
12.	Nitrofurantoin A12-300mcg	20-25	21	18-22	22	--	--	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

Soniya		
TESTED BY (Soniya Sebastian) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com