CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Novobiocin Discs (Symbol: NV; Potency: 30mcg)	
Product Code: AS068	**Expiry Date: 24/12/2019**
Lot Number: 04580616	**Release Date: 24/06/2016**
Analytical Report Number: QAD/AR-16/06-908	**Page No: 1 of 1**

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "NV" on each side of the disc.	Filter paper discs of 6 mm diameter with printed " NV" on each side of the disc.	Complies

Cultural Parameters:

Sr. No.	Test Organisms	Media used (code number)	Incubation time / temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	*Staphylococcus aureus ATCC 25923*	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	22-31	29	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

Soniya		
TESTED BY (Soniya Sebastian) MICROBIOLOGIST	**CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY**	**APPROVED BY (RupaliSabale) HEAD QUALITY ASSURANCE**

End of the Report

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter
Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com