CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Ceftazidime Discs (Symbol: CAZ or CT; Potency: 30 mcg)	
Product Code: AS010	**Expiry Date: 17/04/2020**
Lot Number: 07071016	**Release Date: 17/10/2016**
Analytical Report Number: QAD/AR-16/10-1335	**Page No: 1 of 1**

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "CAZ or CT" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "CAZ "on each side of the disc.	Complies

Cultural Parameters:

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	*Escherichia coli ATCC 25922*	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	25-32	30	Complies
2	*Staphylococcus aureus ATCC 25923*	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	16-20	20	Complies
3	*Pseudomonas aeruginosa ATCC 27853*	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	22-29	28	Complies
4	*Haemophilus influenzae ATCC 49247*	Haemophilus Test Agar Base(DM1273) with added Haemophilus Growth Supplement (MS085)	35-37°C/ 18-24 hrs in 5-7% CO_2	27-35	30	Complies
5	*Neisseria gonorrhoeae ATCC 49226*	GC Agar Base with 1% GC supplement(MS016)	35-37°C/ 18-24 hrs in 5-10% CO_2 + 70% humidity	35-43	38	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

Divya	Poojari	Sabale
TESTED BY (Divya Sasidharan) MICROBIOLOGIST	**CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY**	**APPROVED BY (Rupali sabale) HEAD QUALITY ASSURANCE**

End of the Report

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com