

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name : Thioglycllate Broth Product Code : BC011A Lot Number : 07870919 Analytical Report Number : QAD/AR-19/09-0977				Mfg. Date	. Date 18-Sep-2		
				Expiry Date 22-Sep-2023 Release Date 22-9-2019		22-Sep-2023	
						9	
				Page No	Page 1		
PREINCUBATION C	HECK OE DI ATE						
TESTS CONDUCTED SPECIFICATIONS			OBSERVATIONS			REMAR	KS
Color of Medium		Yellow Color Solution	Yellow Color Solu			Complie	
Clarity of Medium(After Sterilization)		Clear Solution Without Any Precipitate		/ithout Any Precipitate Complies			
pH (After sterilization)		6.9-7.3	7.21			Complie	
Quantity of Medium		70ml	70ml			Complies	
Pre Incubation Check (Negative Control)		No growth should be observed after		vas observed after incubation		Complies	
(-8		incubation at 30-35 °C for 24 - 48 hrs		at 30-35 °C for 24 - 48 hrs		complies	
Туре	Organisms			No. of Organisms	Growth	Growth	
				inoculated			
Standard Salmonella Abony NCTC 6017				50-100	Good lu	xuriant	
Test	Salmonella Abony NCTC 6017			86	Good luxuriant		Complie
Standard				50-100	Good luxuriant		
Test		occus Aureus ATCC 6538		83	Good luxuriant		Complie
Standard	Escherichia	Coli ATCC 8739		50-100	Good lu	xuriant	•
Test	Escherichi	a Coli ATCC 8739		74	Good lu	xuriant	Complie
Standard		nas Aeruginosa ATCC 9027		50-100	Good luxuriant		
Test		nas Aeruginosa ATCC 9027		89	Good lu	xuriant	Complie
Standard	Bacillus Subtilis ATCC 6633			50-100	Good lu	xuriant	
Test		btilis ATCC 6633		75		xuriant	Complie
Standard		n Perfringens ATCC 13124		50-100	Good lu		
Test	Clostridium Perfringens ATCC 13124			56			Complie
Standard Test	, , , , , , , , , , , , , , , , , , , ,			50-100	Good lu		
	Clostridiur	n Sporogenes ATCC 19404		82	Good lu	xuriant	Complie

HEAD MICROBIOLOGY

End of the Report

Madhavi

Rupali

HEAD QUALITY ASSURANCE

Priyanka

MICROBIOLOGIST