



PRODUCT SPECIFICATION SHEET

Soyabean Casein Digest Agar Plate (RP024)

Intended Use

Soyabean Casein Digest Agar Plate (RP024) medium is recommended for cultivation of wide variety of microorganisms and for Sterility testing in pharmaceuticals.

Product Summary and Explanation

In 1955, Leavitt et al.⁽¹⁾ discovered that Soyabean Casein Digest Agar facilitated vigorous growth of aerobic and anaerobic microorganisms. Soyabean Casein Digest Agar is a widely used medium, which supports the growth of wide variety of organisms even that of fastidious ones such as *Neisseria*, *Listeria*, and *Brucella* etc. The medium is used in USP Growth Promotion testing and when testing the suitability of counting methods in the presence of product.⁽²⁾ TSA has a multitude of uses in the clinical laboratory including maintenance of stock cultures, plate counting, isolation of microorganisms from a variety of specimen types and as a base for media containing blood.⁽³⁻⁶⁾ It is also recommended for use in industrial applications when testing water and wastewater,⁽⁷⁾ food,⁽⁸⁻¹³⁾ dairy products,⁽¹⁴⁾ and cosmetics.^(9,15) The medium with addition of blood provides perfectly defined haemolysis zones, while preventing the lysis of erythrocytes due to its sodium chloride content. It has been frequently used in the health industry to produce antigens, toxins etc. Its simple and inhibitor-free composition makes it suitable for the detection of antimicrobial agents in the food and other products. Tryptone Soya Agar is recommended by various pharmacopoeias as sterility testing medium.^(2, 16) Tryptone Soya Agar conforms as per USP⁽²⁾ and is used in microbial limit test and antimicrobial preservative - effective test. Gunn et al.⁽¹⁷⁾ used this medium for the growth of fastidious organisms and study of haemolytic reaction after addition of 5% v/v blood. Soyabean Casein Digest Agar does not contains X and V growth factors. It can be conveniently used in determining the requirements of these growth factors by isolates of *Haemophilus* by the addition of X-factor (ID007), V-factor (ID008), and X+V factor discs (ID009) factor to inoculated TSA plates.⁽⁴⁾

Principles of the Procedure

Soyabean Casein Digest Agar Plate contains pancreatic digest of casein and papaic digest of soyabean which provides amino acids, long chain peptides and essential nutrients required for the growth of microorganisms. Sodium chloride maintains the osmotic balance. *Haemophilus* species may be differentiated by their requirements for X and V factors. Paper strips impregnated with these factors are placed on the surface of the medium after inoculation with the test organism. Following incubation, a zone of growth around the strip indicates a requirement for the factor(s).

Formula / Liter

Ingredients	Gms / Litre
Pancreatic digest of casein	15.00
Papaic digest of soyabean meal	5.00
Sodium chloride	5.00
Agar	15.00

Formula may be adjusted and/or supplemented as required to meet performance specifications

Precautions

1. Prepared plated media are For *in vitro* Diagnostic Use or For Laboratory Use as labeled.
2. Directions for use should be read and followed carefully.
3. If excessive moisture is observed, invert the bottom over an off-set lid and allow to air dry in order to prevent formation of a seal between the top and bottom of the plate during incubation.
4. Observe aseptic techniques and established precautions against microbiological hazards throughout all procedures, since it must be assumed that all specimens/samples collected might contain infectious microorganisms.

Product Deterioration

Do not use plates if they show evidence of microbial contamination, discoloration, drying, cracking or other signs of deterioration.





PRODUCT SPECIFICATION SHEET

Directions

Either streak, inoculate or surface spread the test inoculum (50-100 CFU) aseptically on the plate.

Quality Control Specifications

Appearance	Sterile Soyabean Casein Digest Agar in 90mm plates
Colour	Light yellow coloured medium
Reaction	7.10- 7.50
Quantity of medium	25ml of medium in 90mm plates

Sterility Check: Passes release criteria.

Cultural Response

Recovery rate is considered 100% for bacteria growth on Blood Agar and fungus growth on Sabouraud Dextrose Agar.

Expected Cultural Response: Growth Promotion was observed after an incubation at 30-35°C for 18-24 hours.

Sr. No.	Organisms	Inoculum (CFU)	Results to be achieved CFU)				
			Growth	Observed Lot value (CFU)	Recovery	Incubation Temperature	Incubation period
1.	<i>Bacillus subtilis</i> ATCC 6633	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
2.	<i>Staphylococcus aureus</i> ATCC 25923	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
3.	<i>Staphylococcus aureus</i> ATCC 6538	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
4.	<i>Escherichia coli</i> ATCC 25922	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
5.	<i>Escherichia coli</i> ATCC 8739	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
6.	<i>Escherichia coli</i> NCTC 9002	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
7.	<i>Pseudomonas aeruginosa</i> ATCC 27853	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
8.	<i>Pseudomonas aeruginosa</i> ATCC 9027	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
9.	<i>Salmonella</i> Abony NCTC 6017	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
10.	<i>Micrococcus luteus</i> ATCC 9341	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
11.	<i>Streptococcus pneumonia</i> ATCC 6305	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
12.	<i>Salmonella</i> Typhimurium ATCC 14028	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	18 -24 hrs
13.	<i>Candida albicans</i> ATCC 10231	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	≤5 d
14.	<i>Candida albicans</i> ATCC 2091	50 - 100	luxuriant	35 -100	≥70 %	30 -35 °C	≤5 d
15.	<i>Aspergillus brasiliensis</i> ATCC 16404	50 - 100	good-luxuriant	25 -70	50-70%	30 -35 °C	≤5 d
16.	<i>Aspergillus brasiliensis</i> ATCC 16404	50 - 100	luxuriant	35 -100	≥70 %	20 -25 °C	≤5 d

The organisms listed are the minimum that should be used for quality control testing.

Test Procedure





PRODUCT SPECIFICATION SHEET

Refer appropriate references for standard test procedures.

Results

Refer appropriate references and procedures for interpretation of results.

Storage

On receipt, store plates at 15-25°C.

Expiration

Refer to the expiration date stamped on the pack. Prepared plates stored in their original sleeve wrapping at 15 -25°C until just prior to use may be inoculated up to the expiration date and incubated for recommended incubation times.

Product Disposal

After use, prepared plates, specimen/sample containers and other contaminated materials must be sterilized before discarding.

Limitations of the Procedure

1. Some diagnostic tests may be performed with the primary plate. However, a pure culture is recommended for the majority of biochemical tests and other identification procedures.
2. Consult appropriate references for further information.

Packaging

Product Name : Soyabean Casein Digest Agar Plate

Product Code : RPO24

Available Pack sizes : Pack of 10 plates

References

1. Leavitt, J. M., I. J. Naidorf and P. Shugaevsky. 1955. The undetected anaerobe in endodontics: a sensitive medium for detection of both aerobes and anaerobes. *The NY J. Dentist*. 25:377-382.
2. United States Pharmacopeial Convention, Inc. 2008. The United States pharmacopeia 31/The national formulary 26, Supp. 1, 8-1-08, online. United States Pharmacopeial Convention, Inc., Rockville, Md.
3. MacFaddin. 1985. Media for isolation-cultivation-identification-maintenance of medical bacteria, vol. 1. Williams & Wilkins, Baltimore, Md.
4. Forbes, Sahn and Weissfeld. 2007. Bailey & Scott's diagnostic microbiology, 12th ed. Mosby Inc., St. Louis, Mo.
5. Murray, Baron, Jorgensen, Landry and Pfaller (ed.). 2007. Manual of clinical microbiology, 9th ed. American Society for Microbiology, Washington, D.C.
6. Isenberg and Garcia (ed.). 2004 (update, 2007). Clinical microbiology procedures handbook, 2nd ed. American Society for Microbiology, Washington, D.C.
7. Eaton, Rice and Baird (ed.). 2005. Standard methods for the examination of water and wastewater, 21st ed., online. American Public Health Association, Washington, D.C.
8. Downes and Ito (ed.). 2001. Compendium of methods for the microbiological examination of foods, 4th ed. American Public Health Association, Washington, D.C.
9. U.S. Food and Drug Administration. 2001. Bacteriological analytical manual, online. AOAC International, Gaithersburg, Md.
10. U.S. Department of Agriculture. Microbiology laboratory guidebook, online. Food Safety and Inspection Service, USDA, Washington, D.C.
11. Horwitz (ed.). 2007. Official methods of analysis of AOAC International, 18th ed., online. AOAC International, Gaithersburg, Md.
12. Health Canada. The compendium of analytical methods, online. Food Directorate, Health Products and food Branch, Health Canada, Ottawa, Ontario Canada.
13. International Organization for Standardization. 1994. Microbiology - General guidance for the detection of presumptive pathogenic *Yersinia enterocolitica*. ISO 10273, 1st ed., 1994-12-15. International Organization for Standardization, Geneva, Switzerland.
14. Wehr and Frank (ed.). 2004. Standard methods for the examination of dairy products, 17th ed. American Public Health Association, Washington, D.C.
15. Curry, Joyce and McEwen. 1993. CTFA microbiology guidelines. The Cosmetic, Toiletry and Fragrance Association, Inc., Washington, D.C.
16. Indian Pharmacopoeia, 2007, Govt. of India, Ministry of Health and Family Welfare, New Delhi, India.
17. Gunn B. A., Ohashi D K., Gaydos C. A., Holt E. S., 1977, *J. Clin. Microbiol.*, 5(6) : 650.
- 18.





PRODUCT SPECIFICATION SHEET

Further Information

For further information please contact your local MICROMASTER Representative.



MICROMASTER LABORATORIES PRIVATE LIMITED

RP024PSS, Rev.00, Ver.00

Unit 38/39, Kalpataru Industrial Estate,

Off G.B. Road, Near 'R-Mall' , Thane (W) - 400607. M.S. INDIA.

Ph: +91-9320126789/9833630009/9819991103

Email: sales@micromasterlab.com

Disclaimer :

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

