



PRODUCT SPECIFICATION SHEET

0.1 gm Toluidine Blue (MS151)

Intended Use

0.1 gm Toluidine blue Supplement is recommended for detection of deoxyribonuclease activity.

Composition

Each vial contains the following:

*Ingredients	Concentration
Toluidine Blue	0.1 gm

Directions

1. One vial, sufficient for 1000 ml medium.
2. Rehydrate the contents of 1vial of Toluidine Blue with 5 ml of Distilled water.
3. Mix well and aseptically add to 1000 ml of sterile, molten, cooled (45-50°C) DNase Test Agar Base (DM075).
4. Mix well and pour into sterile petri plates.

Storage and Shelf Life

Store at 2 - 8°C. Use before expiry date on the label.

Note: * Not for medicinal use

Further Information

For further information please contact your local MICROMASTER Representative.



MICROMASTER LABORATORIES PRIVATE LIMITED

Unit 38/39, Kalpataru Industrial Estate,
Off G.B. Road, Near 'R-Mall' , Thane (W) - 400607. M.S. INDIA.
Ph: +91-9320126789/9833630009/9819991103
Email: sales@micromasterlab.com

MS151PSS, QAD/FR/024, Rev.00

Disclaimer:

All Products conform exclusively to the information contained in this and other related Micromaster Publications. Users must ensure that the product(s) is appropriate for their application, prior to use. The information published in this publication is based on research and development work carried out in our laboratory and is to the best of our knowledge true and accurate. Micromaster Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are intended for laboratory, diagnostic, research or further manufacturing use only and not for human or animal or therapeutic use, unless otherwise specified. Statements included herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

