






## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

<b>Product Name: Fouchet's Reagent</b>	
<b>Product Code: AR022</b>	<b>Expiry Date : 27/04/2026</b>
<b>Lot Number: 02460418</b>	<b>Release Date: 26/04/2018</b>
<b>Analytical Report Number: QAD/AR-18/04-393</b>	<b>Page No : Page 1 of 1</b>

TESTS CONDUCTED	SPECIFICATIONS		OBSERVATIONS		REMARKS
Appearance	Yellow colored solution		Yellow colored solution		Complies
Quality Control testing for presences of bile pigment	Presence of bile pigment	Green color	Presence of bile pigment	Green color	Complies
	Absence of bile pigment	colorless	Absence of bile pigment	colorless	
Clarity	Clear solution		Clear solution		Complies
Use	Used to detect bile pigment in urine samples				

### Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY (Madhura Kadam) MICROBIOLOGIST</b>	<b>CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY</b>	<b>APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE</b>

End Of The Report