

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Amikacin Discs (Symbol: AK or AM; Potency: 30mcg)	
Product Code: AS001	Expiry Date: 20/07/2020
Lot Number: 00410117	Release Date: 20/01/2017
Analytical Report Number: QAD/AR-17/01-149	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "AK or AM" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "AK " on each side of the discs	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used (Product Code)	Incubation time and temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	19-26	26mm	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	20-26	23 mm	Complies
3	<i>Pseudomonas aeruginosa</i> ATCC 27853	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	18-26	21 mm	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Divya Sasidharan) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
Ph: +91-22-2589 5505, 4760, 4681, 9320126789
Email: micromaster@micromasterlab.com

