CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Amikacin Discs (Symbol: AK or AM; Potency: 30mcg)	
Product Code: AS001	**Expiry Date: 21/06/2021**
Lot Number: 09541217	**Release Date: 21/12/2017**
Analytical Report Number: QAD/AR-17/12-1478	**Page No: 1 of 1**

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "AK or AM" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "AK or AM" on each side of the discs	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used (Product Code)	Incubation time and temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	*Escherichia coli ATCC 25922*	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	19-26	26	Complies
2	*Staphylococcus aureus ATCC 25923*	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	20-26	26	Complies
3	*Pseudomonas aeruginosa ATCC 27853*	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	18-26	26	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

Kadam	Modak	Sabale
TESTED BY ((Madhura Kadam) MICROBIOLOGIST	**CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY**	**APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE**

End of the Report

Micromaster Laboratories Pvt. Ltd

Manufacturer & Exporter

Unit No.38/39, Kalpataru Industrial Estate, Near Runwal Estate, Behind R Mall, Ghodbunder Road, Thane (W) - 400607, Maharashtra, India

Tel : +91-22-2589 5505 / 2589 4681 / 2589 4760 sales@micromasterlab.com