



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Amoxicillin Discs (Symbol: AMX; Potency: 10mcg)	
Product Code: AS002	Expiry Date: 30/04/2022
Lot Number: 06991018	Release Date: 31/10/2018
Analytical Report Number: QAD/AR-18/09-1067	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "AMX" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "AMX" on each side of the discs	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used (Product Code)	Incubation time and temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	19-25	24	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	28-36	33	Complies

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Pooja Sutar) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

AS002COA,QAD/FR/031,Rev.00/01.01.2018

Micromaster laboratories Pvt.Ltd
Manufacturer and Exporter
Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,
Ghodbunder Road,Thane (W)-400607,Maharashtra,India
[Tel:- +91-22-2589 5505](tel:+91-22-25895505)sales@micromasterlab.com