



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Ampicillin Discs (Symbol: AMP ; Potency: 10mcg)	
Product Code: AS003	Expiry Date: 01/06/2020
Lot Number: 08711116	Release Date: 01/12/2016
Analytical Report Number: QAD/AR-16/11-1552	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "AMP " on each side of the disc.	Filter paper discs of 6 mm diameter with printed " AMP " on each side of the disc.	Complies

### Cultural Parameters:

Sr. No.	Test Organisms	Media used (code number)	Incubation time / temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	16-22	18	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	27-35	32	Complies
3	<i>Haemophilus influenza</i> ATCC 49247	Haemophilus Test Agar Base(DM1273) with added Haemophilus Growth Supplement (MS085)	35-37°C/ 18-24 hrs in 5-7% CO <sub>2</sub>	13-21	20	Complies
4	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	30-36	34	Complies

### Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Divya Sasidharan) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.  
Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kaipataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,  
Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.  
Ph: +91-22-2589 5505, 4760, 4681, 9320126789  
Email: micromaster@micromasterlab.com

