



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Amoxicillin +Clavulanic acid Discs (Symbol: AMC or AU; Potency: 20+10mcg)	
Product Code: AS004	Expiry Date: 09/12/2021
Lot Number: 03400618	Release Date: 13/06/2018
Analytical Report Number: QAD/AR-18/06-546	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "AMC or AU" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "AMC or AU" on each side of the disc.	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	18-24	24	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	28-36	32	Complies
3	<i>Haemophilus influenzae</i> ATCC 49247	Haemophilus Test Agar Base(DM1273) with added Haemophilus Growth Supplement (MS085)	35-37°C/ 18-24 hrs in 5-7% CO ₂	15-23	20	Complies
4	<i>Escherichia coli</i> ATCC 35218	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	17-22	19	Complies

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

AS004COA,QAD/FR/031,Rev.00/01.01.2018

Micromaster laboratories Pvt.Ltd

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,
Ghodbunder Road ,Thane (W)-400607,Maharashtra,India

[Tel:- +91-22-2589 5505](tel:+91-22-25895505)sales@micromasterlab.com