



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

| | |
|---|--------------------------|
| Product Name: Cefazolin Discs (Symbol: CZ; Potency: 30 mcg) | |
| Product Code: AS008 | Expiry Date: 07/03/2022 |
| Lot Number: 05410918 | Release Date: 07/09/2018 |
| Analytical Report Number: QAD/AR-18/09-887 | Page No: 1 of 1 |




| TEST CONDUCTED | SPECIFICATION | OBSERVATIONS | REMARKS |
|----------------|---|---|----------|
| Appearance | Filter paper discs of 6 mm diameter with printed "CZ" on each side of the disc. | Filter paper discs of 6 mm diameter with printed "CZ" on each side of the disc. | Complies |

Cultural Parameters:

| Sr. No. | Test Organisms | Media used (code number) | Incubation time and temperature | Standard Zone Diameter | Zone Diameter Observed In mm | Remarks |
|---------|--|-----------------------------|---------------------------------|------------------------|------------------------------|----------|
| 1 | <i>Escherichia coli</i> ATCC 25922 | Mueller Hinton Agar (DM172) | 32-35°C/ 18-24 hrs | 21-27 | 25 | Complies |
| 2 | <i>Staphylococcus aureus</i> ATCC 25923 | Mueller Hinton Agar (DM172) | 32-35°C/ 18-24 hrs | 29-35 | 33 | Complies |

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

| | | |
|---|--|---|
|  |  |  |
| TESTED BY (Madhura Kadam) MICROBIOLOGIST | CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY | APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE |

End of the Report