



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Ciprofloxacin Discs (Symbol: CIP or CP; Potency: 5 mcg)	
Product Code: AS018	Expiry Date: 11/01/2022
Lot Number: 03990718	Release Date: 11/07/2018
Analytical Report Number: QAD/AR-18/07-603	Page No: 1 of 2

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "CIP or CP" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "CIP or CP" on each side of the disc.	Complies

Cultural Parameters:

Sr. No.	Test Organisms	Media used with (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	30-40	32	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	22-30	22	Complies
3	<i>Pseudomonas aeruginosa</i> ATCC 27853	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	25-33	27	Complies
4	<i>Haemophilus influenzae</i> ATCC 49247	Haemophilus Test Agar Base(DM1273) with added Haemophilus Growth Supplement (MS085)	35-37°C/ 18-24 hrs in 5-7% CO ₂	34-42	38	Complies
5	<i>Neisseria gonorrhoeae</i> ATCC 49226	GC Agar Base with 1% GC supplement(MS016)	35-37°C/ 18-24 hrs in 5-10% CO ₂ + 70% humidity	48-58	49	Complies

AS018COA,QAD/FR/031,Rev.00/01.01.2018

Micromaster laboratories Pvt.Ltd

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,
Ghodbunder Road ,Thane (W)-400607,Maharashtra,India

[Tel:- +91-22-2589 5505sales@micromasterlab.com](mailto:+91-22-25895505sales@micromasterlab.com)






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Ciprofloxacin Discs (Symbol: CIP or CP; Potency: 5 mcg)	
Product Code: AS018	Expiry Date: 11/01/2022
Lot Number: 03990718	Release Date: 11/07/2018
Analytical Report Number: QAD/AR-18/07-603	Page No: 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

AS018COA,QAD/FR/031,Rev.00/01.01.2018

Micromaster laboratories Pvt.Ltd

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,
Ghodbunder Road ,Thane (W)-400607,Maharashtra,India

Tel:- +91-22-2589 5505sales@micromasterlab.com