



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Clindamycin Discs (Symbol: CD or CY; Potency: 2mcg)	
Product Code: AS019	Expiry Date: 15/09/2021
Lot Number: 01540318	Release Date: 14/02/2018
Analytical Report Number: QAD/AR-18/03-233	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "CD or CY " on each side of the disc.	Filter paper discs of 6 mm diameter with printed " CD or CY " on each side of the disc.	Complies

### Cultural Parameters:.

Sr. No.	Test Organisms	Media used (code number)	Incubation time / temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	24-30	27	Complies
4	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	19-25	21	Complies

### Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Madhura Kadam) <b>MICROBIOLOGIST</b>	<b>CHECKED BY</b> (Aditi Modak) <b>HEAD MICROBIOLOGY</b>	<b>APPROVED BY</b> (Rupali Sabale) <b>HEAD QUALITY ASSURANCE</b>

End of the Report