



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Furazolidone Discs (Symbol: F; Potency: 50mcg)	
Product Code: AS025-1v1	Expiry Date: 11/01/2022
Lot Number: 04020718	Release Date: 11/07/2018
Analytical Report Number: QAD/AR-18/07-600	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "F" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "F" on each side of the disc.	Complies

### Cultural Parameters:.

Sr. No.	Test Organisms	Media used (code number)	Incubation time / temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	20-25	22	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	18-22	20	Complies

### Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report