



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Kanamycin Discs (Symbol: K; Potency: 30mcg)	
Product Code: AS027-1vl	Expiry Date: 11/10/2021
Lot Number: 02200418	Release Date: 06/04/2014
Analytical Report Number: QAD/AR-18/04-353	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "K" on each side of the disc.	Filter paper discs of 6 mm diameter with printed " K " on each side of the disc.	Complies

Cultural Parameters:.

Sr. No.	Test Organisms	Media used (code number)	Incubation time / temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	17-25	25	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	19-26	26	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Deepti Nair) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report