

CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY



Product Name: Norfloxacin Discs (Symbol: NO or NX ; Potency: 10 mcg)	
Product Code: AS035-1VI	Expiry Date: 22/04/2020
Lot Number: 07291016	Release Date: 22/10/2016
Analytical Report Number: QAD/AR-16/10-1370	Page No: 1 of 1

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "NO" or "NX" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "NO" or "NX" on each side of the disc.	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	28-35	35	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar DM172)	32-35°C/ 18-24 hrs	17-28	28	Complies
3	<i>Pseudomonas aeruginosa</i> ATCC 27853	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	22-29	29	Complies
4	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	15-21	20	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Divya Sasidharan) MICROBIOLOGIST	CHECKED BY (Vijayalakshmi Poojari) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Result

Rev.00

10/12/2014

QAD/FR/01

Micromaster Laboratories Pvt. Ltd.
 Manufacture & Exporter

Culturing your needs in Microbiology...

38/39, Kalpataru Industrial Estate, Opp Lawkim, Off. Ghodbunder Rd.,
 Near Runwal Estate, Thane (W) 400 607, Maharashtra, INDIA.
 Ph: +91-22-2589 5505, 4760, 4681, 9320126789
 Email: micromaster@micromasterlab.com

