



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Norfloxacin Discs (Symbol: NO or NX ; Potency: 10 mcg)	
Product Code: AS035-1VI	Expiry Date: 26/10/2021
Lot Number: 02340418	Release Date: 26/04/2018
Analytical Report Number: QAD/AR-18/04-388	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "NO" or "NX" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "NO" or "NX" on each side of the disc.	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	28-35	32	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar DM172)	32-35°C/ 18-24 hrs	17-28	27	Complies
3	<i>Pseudomonas aeruginosa</i> ATCC 27853	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	22-29	27	Complies
4	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	15-21	21	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Deepti Nair) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Result