



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Ofloxacin Discs (Symbol: OF; Potency: 5 mcg)	
Product Code: AS036	Expiry Date: 07/03/2022
Lot Number: 05440918	Release Date: 07/09/2018
Analytical Report Number: QAD/AR-18/09-885	Page No: 1 of 2

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "OF" on each side of the disc.	Filter paper discs of 6 mm diameter with printed " OF " on each side of the disc.	Complies

Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	29-33	32	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar DM172)	32-35°C/ 18-24 hrs	24-28	26	Complies
3	<i>Pseudomonas aeruginosa</i> ATCC 27853	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	17-21	20	Complies
4	<i>Haemophilus influenzae</i> ATCC 49247	Haemophilus Test Agar Base(DM1273) with added Haemophilus Growth Supplement (MS085)	35-37°C/ 18-24 hrs in 5-7% CO ₂	31-40	35	Complies
5	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	16-21	19	Complies
6	<i>Neisseria gonorrhoeae</i> ATCC 49226	GC Agar Base with 1% GC supplement(MS016)	35-37°C/ 18-24 hrs in 5-10% CO ₂ + 70% humidity	43-51	43	Complies






CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Ofloxacin Discs (Symbol: OF; Potency: 5 mcg)	
Product Code: AS036	Expiry Date: 07/03/2022
Lot Number: 05440918	Release Date: 07/09/2018
Analytical Report Number: QAD/AR-18/09-885	Page No: 2 of 2

Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report