



CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Rifampicin Discs (Symbol: R or RIF; Potency: 5mcg)	
Product Code: AS041-1v1	Expiry Date: 13/12/2021
Lot Number: 03500618	Release Date: 13/06/2018
Analytical Report Number: QAD/AR-18/06-545	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "R or RIF " on each side of the disc.	Filter paper discs of 6 mm diameter with printed " R or RIF " on each side of the disc.	Complies

Cultural Parameters:.

Sr. No.	Test Organisms	Media used (code number)	Incubation time / temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	08-10	10	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	26-34	32	Complies
3	<i>Haemophilus influenza</i> ATCC 49247	Haemophilus Test Agar Base(DM1273) with added Haemophilus Growth Supplement (MS085)	35-37°C/ 18-24 hrs in 5-7% CO ₂	22-30	26	Complies
4	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	25-30	28	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
TESTED BY (Madhura Kadam) MICROBIOLOGIST	CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY	APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report

AS041COA,QAD/FR/031,Rev.00/01.01.2018

Micromaster laboratories Pvt.Ltd

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,
Ghodbunder Road ,Thane (W)-400607,Maharashtra,India

Tel:- +91-22-2589 5505sales@micromasterlab.com