



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Cefpodoxime Discs (Symbol:CPD; Potency: 10mcg)	
Product Code: AS062	Expiry Date: 11/01/2022
Lot Number: 03980718	Release Date: 12/07/2018
Analytical Report Number: QAD/AR-18/07-611	Page No: 1 of 1




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "CPD" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "CEP" on each side of the disc.	Complies

### Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	23-28	25	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	19-25	23	Complies

### Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Madhura Kadam) MICROBIOLOGIST	<b>CHECKED BY</b> (Aditi Modak) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report