



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

<b>Product Name: Cefipime Discs (Symbol: CPM; Potency: 30 mcg)</b>	
<b>Product Code: AS066</b>	<b>Expiry Date: 07/03/2022</b>
<b>Lot Number: 05420918</b>	<b>Release Date: 07/09/2018</b>
<b>Analytical Report Number: QAD/AR-18/09-886</b>	<b>Page No: 1 of 2</b>

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "CPM" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "CPM" on each side of the disc.	Complies

### Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	31-37	35	Complies
2	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	23-29	27	Complies
3	<i>Pseudomonas aeruginosa</i> ATCC 27853	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	24-30	28	Complies
4	<i>Haemophilus influenzae</i> ATCC 49247	Haemophilus Test Agar Base(DM1273) with added Haemophilus Growth Supplement (MS085)	35-37°C/ 18-24 hrs in 5-7% CO <sub>2</sub>	25-31	31	Complies
5	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	28-35	35	Complies
6	<i>Neisseria gonorrhoeae</i> ATCC 49226	GC Agar Base with 1% GC supplement (MS016)	35-37°C/ 18-24 hrs in 5-10% CO <sub>2</sub> + 70% humidity	37-46	37	Complies






## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Cefipime Discs (Symbol: CPM; Potency: 30 mcg)	
Product Code: AS066	Expiry Date: 07/03/2022
Lot Number: 05420918	Release Date: 07/09/2018
Analytical Report Number: QAD/AR-18/09-886	Page No: 2 of 2

### Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Madhura Kadam) MICROBIOLOGIST	<b>CHECKED BY</b> (Aditi Modak) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End of the Report