



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

<b>Product Name: Fluconazole Discs (Symbol: FLC or FU; Potency: 25mcg)</b>	
<b>Product Code: AS075</b>	<b>Expiry Date:12/08/2021</b>
<b>Lot Number: 00810218</b>	<b>Release Date: 12/02/2018</b>
<b>Analytical Report Number: QAD/AR-18/02-117</b>	<b>Page No: 1 of 2</b>

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "FLC or FU" on each side of the disc.	Filter paper discs of 6 mm diameter with printed " FLC or FU" on each side of the disc.	Complies

### Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Candida albicans ATCC 90028</i>	Mueller Hinton Agar ( DM172)+2% Glucose and 0.5%µg/ml Methylene Blue	20-25°C/ 20-48 hrs	27-38	35	Complies
2	<i>Candida parapsilosis ATCC 22019</i>	Mueller Hinton Agar ( DM172)+2% Glucose and 0.5%µg/ml Methylene Blue	20-25°C/ 20-48 hrs	22-33	30	Complies
3	<i>Candida tropicalis ATCC 750</i>	Mueller Hinton Agar ( DM172)+2% Glucose and 0.5%µg/ml Methylene Blue	20-25°C/ 20-48 hrs	16-25	25	Complies
4	<i>Candida albicans ATCC 10231</i>	Mueller Hinton Agar ( DM172)+2% Glucose and 0.5%µg/ml Methylene Blue	20-25°C/ 20-48 hrs	18-22	22	Complies






## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Fluconazole Discs (Symbol: FLC or FU; Potency: 25mcg)	
Product Code: AS075	Expiry Date:12/08/2021
Lot Number: 00810218	Release Date: 12/02/2018
Analytical Report Number: QAD/AR-18/02-117	Page No: 2 of 2

### Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Deepti Nair) <b>MICROBIOLOGIST</b>	<b>CHECKED BY</b> (Aditi Modak) <b>HEAD MICROBIOLOGY</b>	<b>APPROVED BY</b> (Rupali Sabale) <b>HEAD QUALITY ASSURANCE</b>

End of the Report