



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

<b>Product Name: Moxifloxacin Discs (Symbol:MO; Potency: 5 mcg)</b>	
<b>Product Code: AS078</b>	<b>Expiry Date: 11/01/2022</b>
<b>Lot Number: 04010718</b>	<b>Release Date: 11/07/2018</b>
<b>Analytical Report Number: QAD/AR-18/07-597</b>	<b>Page No: 1 of 2</b>




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "MO" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "MO" on each side of the disc.	Complies

### Cultural Parameters

Sr. No.	Test Organisms	Media used (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Escherichia coli</i> ATCC 25922	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	28-35	28	Complies
2	<i>Pseudomonas aeruginosa</i> ATCC 27853	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	17-25	25	Complies
3	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	28-35	31	
4	<i>Haemophilus influenzae</i> ATCC 49247	Haemophilus Test Agar Base(DM1273) with added Haemophilus Growth Supplement (MS085)	35-37°C/ 18-24 hrs in 5-7% CO <sub>2</sub>	31-39	32	Complies
5	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	25-31	31	Complies

### Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY (Madhura Kadam) MICROBIOLOGIST</b>	<b>CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY</b>	<b>APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE</b>

End of the Report



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

<b>Product Name: Moxifloxacin Discs (Symbol:MO; Potency: 5 mcg)</b>	
<b>Product Code: AS078</b>	<b>Expiry Date: 11/01/2022</b>
<b>Lot Number: 04010718</b>	<b>Release Date: 11/07/2018</b>
<b>Analytical Report Number: QAD/AR-18/07-597</b>	<b>Page No: 2 of 2</b>