



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

<b>Product Name: Oxacillin Discs OX 1mcg (Symbol: OX; Potency: 1 mcg)</b>	
<b>Product Code: AS091</b>	<b>Expiry Date: 15/09/2021</b>
<b>Lot Number: 01510318</b>	<b>Release Date: 14/02/2018</b>
<b>Analytical Report Number: QAD/AR-18/03-230</b>	<b>Page No: 1 of 1</b>




TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed "OX" on each side of the disc.	Filter paper discs of 6 mm diameter with printed "OX" on each side of the disc.	Complies

### Cultural Parameters

Sr. No.	Test Organisms	Media used with (code number)	Incubation time and temperature	Standard Zone Diameter In mm	Zone Diameter Observed In mm	Remarks
1	<i>Staphylococcus aureus</i> ATCC 25923	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	18-24	20mm	Complies
2	<i>Streptococcus pneumoniae</i> ATCC 49619	Mueller Hinton Agar (DM172) with 5% sheep blood	32-35°C/ 40-48 hrs	≤12	8mm	Complies

### Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY (Madhura Kadam) MICROBIOLOGIST</b>	<b>CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY</b>	<b>APPROVED BY (Rupali Sabale) HEAD QUALITY ASSURANCE</b>

End of the Report