CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: Phosphomycin	
Product Code: AS132	**Expiry Date: 14/05/2020**
Lot Number: 08021117	**Release Date: 14/11/2017**
Analytical Report Number: QAD/AR-17/11-1257	**Page No: 1 of 1**

TEST CONDUCTED	SPECIFICATION	OBSERVATIONS	REMARKS
Appearance	Filter paper discs of 6 mm diameter with printed " CFS " on each side of the disc.	Filter paper discs of 6 mm diameter with printed " CFS " on each side of the disc.	Complies

Cultural Parameters:

Sr. No.	Test Organisms	Media used (code number)	Incubation time / temperature	Standard Zone Diameter	Zone Diameter Observed In mm	Remarks
1	*Escherichia coli ATCC 25922*	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	22-30	35mm	Complies
2	*Staphylococcus aureus ATCC 25923*	Mueller Hinton Agar (DM172)	32-35°C/ 18-24 hrs	25-33	28mm	Complies

Final Results: COMPLIES

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

TESTED BY (Madhura Kadam) MICROBIOLOGIST	**CHECKED BY (Aditi Modak) HEAD MICROBIOLOGY**	**APPROVED BY (Rupale Sabale) HEAD QUALITY ASSURANCE**

End of the Report

Micromaster Laboratories Pvt. Ltd
Manufacturer & Exporter
Unit No.38/39, Kalpataru Industrial Estate, Near Runwal Estate, Behind R Mall,
Ghodbunder Road, Thane (W) - 400607, Maharashtra, India
Tel : +91-22-2589 5505 / 2589 4681 / 2589 4760 sales@micromasterlab.com