



## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

<b>Product Name: MR-VP Broth (Glucose phosphate broth)</b>	
<b>Product Code: BL008-5ml-24 BL</b>	<b>Expiry Date: 21/07/2022</b>
<b>Lot Number: 04290718</b>	<b>Release Date: 26/07/2018</b>
<b>Analytical Report Number: QAD/AR-18/07-684</b>	<b>Page No: 1 of 2</b>

<b>PREPARED MEDIUM</b>			
<b>TEST CONDUCTED</b>	<b>SPECIFICATION</b>	<b>OBSERVATIONS</b>	<b>REMARKS</b>
Color of medium	Light yellow colored medium	Light yellow colored medium	Complies
Clarity of medium (After sterilization)	Clear solution without any precipitate	Clear solution without any precipitate	Complies
pH (After sterilization)	6.70 -7.10 at 25°C	6.70	Complies
Gel strength	Not Applicable	Not Applicable	Not Applicable
Quantity of Medium	5 ml	5ml	Complies
<b>PREINCUBATION CHECK OF MEDIUM</b>			
Pre Incubation Check	No growth should be observed after incubation at 30-35 °C for 24 - 48 hrs	No growth observed after incubation at 30-35 °C for 24 - 48 hrs	Complies

### **CULTURAL PARAMETERS:**

<b>Organisms</b>	<b>Incubation time/ temp</b>	<b>No of organisms inoculated</b>	<b>Growth</b>	<b>MR Test</b>	<b>VPTest</b>	<b>Remarks</b>
<i>Enterobacter aerogenes</i> ATCC 13048	30-32°C / 18-48 hrs.	75	Luxuriant	Negative reaction	Positive reaction, eosin pink/red color within 2-5 minutes	Complies
<i>Escherichia coli</i> ATCC 25922	30-32°C / 18-48 hrs.	91	Luxuriant	Positive reaction, bright red color	Negative reaction	Complies
<i>Klebsiella pneumoniae</i> ATCC 23357	30-32°C / 18-48 hrs.	72	Luxuriant	Negative reaction	Positive reaction, eosin pink/red color within 2-5 minutes	Complies

BL008COA,QAD/FR/031,Rev.00/01.01.2018

### **Micromaster laboratories Pvt.Ltd**

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate, Behind R Mall,  
Ghodbunder Road, Thane (W)-400607, Maharashtra, India

Tel:- +91-22-2589 5505sales@micromasterlab.com






## CERTIFICATE OF ANALYSIS, QUALITY AND CONFORMITY

Product Name: MR-VP Broth (Glucose phosphate broth)	
Product Code: BL008-5ml-24 BL	Expiry Date: 21/07/2022
Lot Number: 04290718	Release Date: 26/07/2018
Analytical Report Number: QAD/AR-18/07-684	Page No: 2 of 2

### Final Results: **COMPLIES**

This is to certify that this lot passes and it conforms to the above mentioned tests and specifications.

		
<b>TESTED BY</b> (Madhura Kadam) MICROBIOLOGIST	<b>CHECKED BY</b> (Aditi Modak) HEAD MICROBIOLOGY	<b>APPROVED BY</b> (Rupali Sabale) HEAD QUALITY ASSURANCE

End Of The Report

BL008COA,QAD/FR/031,Rev.00/01.01.2018

#### **Micromaster laboratories Pvt.Ltd**

Manufacturer and Exporter

Unit No.38/39, Kalpataru Industrial Estate .Near Runwal Estate,Behind R Mall,  
Ghodbunder Road,Thane (W)-400607,Maharashtra,India

[Tel:- +91-22-2589 5505](tel:+91-22-25895505)[sales@micromasterlab.com](mailto:sales@micromasterlab.com)